

- Limit seating functions in the User Menu

It is possible to have the amount of seating functions (tilt, lift, recline, etc.) that the user wishes to control on the User Menu. For example, you can have just one seating function on the User Menu.

Below, there is an example of the settings when all the seating actuators are programmed.

Seating						
ISM-8/CxSM						
Intelligent Seating Module						
Global						
Axis Setup	Axis 1	Axis 2	Axis 3	Axis 4	Axis 5	Axis 6
Axis Channels	[ 4 ]	[ 5 ]	[ 3 ]	[ 2 ]	[ 1 ]	[ 12 ]
Axis Display	[ 123 ]	[ 3 6 ]	[ 1 ]	[ 5 ]	[ 4 ]	[ 45 ]
Invert Axis Direction	Yes	Yes	Yes	Yes	Yes	Yes
Axis Name	Tilt	Lift	Recline	Left legrest	Right legrest	Both legrests
Channels	Channel 1	Channel 2	Channel 3	Channel 4	Channel 5	Channel 6

In the picture below, you can see the settings when only one seating function is desired to be controlled (Only tilt control in this case).

Seating						
ISM-8/CxSM						
Intelligent Seating Module						
Global						
Axis Setup	Axis 1	Axis 2	Axis 3	Axis 4	Axis 5	Axis 6
Axis Channels	[ 4 ]	[ ]	[ ]	[ ]	[ ]	[ ]
Axis Display	[ 123 ]	[ ]	[ ]	[ ]	[ ]	[ ]
Invert Axis Direction	Yes	Yes	Yes	Yes	Yes	Yes
Axis Name	Tilt	Lift	Recline	Left legrest	Right legrest	Both legrests
Channels	Channel 1	Channel 2	Channel 3	Channel 4	Channel 5	Channel 6

**Warning:** This setting will prevent you from controlling the other seating functions from the joystick. This is due to the fact that seating functions are declared globally on the Rnet.